

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0043
Location:	CHERRY HILL TWSP. P.D.	
Calibration File No.:	02695	Calib. Date: 04/26/2022
Certification File No.:	02696	Calib. No.: 00043
Linearity File No.:	02697	Cert. Date: 04/26/2022
Solution File No.:	02747	Cert. No.: 00039
Sequential File No.:	02747	Lin. Date: 04/26/2022
		Lin. No.: 00040
		Soln. Date: 09/06/2022
		Soln. No.: 00373
		File Date: 09/06/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0007
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0144

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:15D		
Control 1 EC	0.100%	08:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16D		
Control 2 EC	0.100%	08:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17D		
Control 3 EC	0.099%	08:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	08:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WEIST

First Name: JAMES

MI: R

Signature: _____

Sgt JWA 352

Badge No.: 352

Date: 09/06/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0043
Location:	CHERRY HILL TWSP. P.D.	
Calibration File No.:	02695	Calib. Date: 04/26/2022
Certification File No.:	02696	Calib. No.: 00043
Linearity File No.:	02697	Cert. Date: 04/26/2022
Solution File No.:	02735	Cert. No.: 00039
Sequential File No.:	02735	Lin. Date: 04/26/2022
		Lin. No.: 00040
		Soln. Date: 08/08/2022
		Soln. No.: 00372
		File Date: 08/08/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0007
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0090

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:21D		
Control 1 EC	0.101%	12:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22D		
Control 2 EC	0.100%	12:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24D		
Control 3 EC	0.101%	12:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WEIST

First Name: JAMES

MI: R

Signature: _____

S₆T J_{WT} 352

Badge No.: 352

Date: 08/08/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0043
Location:	CHERRY HILL TWSP. P.D.	
Calibration File No.:	02695	Calib. Date: 04/26/2022
Certification File No.:	02696	Cert. No.: 00043
Linearity File No.:	02697	Cert. Date: 04/26/2022
Solution File No.:	02729	Lin. No.: 00040
Sequential File No.:	02729	Lin. Date: 04/26/2022
		Soln. No.: 00371
		Soln. Date: 07/14/2022
		File Date: 07/14/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0007
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0016

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:07D		
Control 1 EC	0.100%	14:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:08D		
Control 2 EC	0.101%	14:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	14:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:10D		
Control 3 EC	0.101%	14:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WEIST

First Name: JAMES

MI: R

Signature: _____

Badge No.: 352

Date: 07/14/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0043
Location:	CHERRY HILL TWSP. P.D.		
Calibration File No.:	02695	Calib. Date:	04/26/2022
Certification File No.:	02696	Cert. Date:	04/26/2022
Linearity File No.:	02697	Lin. Date:	04/26/2022
Solution File No.:	02725	Soln. Date:	06/22/2022
Sequential File No.:	02725	File Date:	06/22/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0007
Solution Control Lot:	21090	Expires:	03/17/2023
		Bottle No.:	0208

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:54D		
Control 1 EC	0.100%	10:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56D		
Control 2 EC	0.100%	10:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57D		
Control 3 EC	0.101%	10:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WEIST

First Name: JAMES

MI: R

Signature: _____

Sgt J [Signature] 352

Badge No.: 352

Date: 06/22/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0043
Location:	CHERRY HILL TWSP. P.D.		
Calibration File No.:	02695	Calib. Date:	04/26/2022
Certification File No.:	02696	Cert. Date:	04/26/2022
Linearity File No.:	02697	Lin. Date:	04/26/2022
Solution File No.:	02714	Soln. Date:	05/23/2022
Sequential File No.:	02714	File Date:	05/23/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0007
Solution Control Lot:	21090	Expires:	03/17/2023
		Bottle No.:	0092

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:34D		
Control 1 EC	0.102%	07:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:35D		
Control 2 EC	0.101%	07:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	07:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:37D		
Control 3 EC	0.100%	07:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	07:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WEIST

First Name: JAMES

MI: R

Signature: _____

Badge No.: 352

Date: 05/23/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARUM-0043
Location:	CHERRY HILL TWSP. P.D.		
Calibration File No.:	02633	Calib. Date:	11/16/2021
Certification File No.:	02634	Cert. No.:	00042
Linearity File No.:	02635	Cert. Date:	11/16/2021
Solution File No.:	02689	Lin. No.:	00039
Sequential File No.:	02689	Lin. Date:	11/16/2021
		Soln. Date:	04/14/2022
		Soln. No.:	00367
		File Date:	04/14/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWA S3-0008
Solution Control Lot:	21270	Expires:	07/21/2023
		Bottle No.:	0323

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:34D		
Control 1 EC	0.101%	13:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35D		
Control 2 EC	0.101%	13:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36D		
Control 3 EC	0.100%	13:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WEIST

First Name: JAMES

MI: R

Signature: _____

Badge No.: 352

Date: 04/14/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0043
Location:	CHERRY HILL TWSP. P.D.	
Calibration File No.:	02633	Calib. Date: 11/16/2021
Certification File No.:	02634	Calib. No.: 00042
Linearity File No.:	02635	Cert. Date: 11/16/2021
Solution File No.:	02679	Cert. No.: 00038
Sequential File No.:	02679	Lin. Date: 11/16/2021
		Lin. No.: 00039
		Soln. Date: 03/14/2022
		Soln. No.: 00366
		File Date: 03/14/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0008
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0088

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:23S		
Control 1 EC	0.101%	07:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	07:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25S		
Control 2 EC	0.101%	07:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26S		
Control 3 EC	0.100%	07:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	07:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WEIST

First Name: JAMES

MI: R

Signature: _____

Sgt JAS 352

Badge No.: 352

Date: 03/14/2022

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARUM-0043
Location:	CHERRY HILL TWSP. P.D.	
Calibration File No.:	02633	Calib. Date: 11/16/2021
Certification File No.:	02634	Calib. No.: 00042
Linearity File No.:	02635	Cert. Date: 11/16/2021
Solution File No.:	02669	Cert. No.: 00038
Sequential File No.:	02669	Lin. Date: 11/16/2021
		Lin. No.: 00039
		Soln. Date: 02/24/2022
		Soln. No.: 00365
		File Date: 02/24/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWA S3-0008
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0089

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:51S		
Control 1 EC	0.100%	08:52S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	08:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53S		
Control 2 EC	0.100%	08:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54S		
Control 3 EC	0.100%	08:55S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WEIST

First Name: JAMES

MI: R

Signature: _____

Sot JWA 352

Badge No.: 352

Date: 02/24/2022

Calibrating Unit

New Standard Solution Report

Equipment

Location: Alcotest 7110 MKIII-C
 CHERRY HILL TWSP. P.D.

Calibration File No.:	02633	Calib. Date:	11/16/2021	Calib. No.:	00042
Certification File No.:	02634	Cert. Date:	11/16/2021	Cert. No.:	00038
Linearity File No.:	02635	Lin. Date:	11/16/2021	Lin. No.:	00039
Solution File No.:	02659	Soln. Date:	01/28/2022	Soln. No.:	00364
Sequential File No.:	02659	File Date:	01/28/2022		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWA S3-0008
Control Solution %:	0.100%			Expires:	03/17/2023
Solution Control Lot:	21090			Bottle No.:	0030

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:09S		
Control 1 EC	0.102%	08:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:10S		
Control 2 EC	0.102%	08:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	08:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:12S		
Control 3 EC	0.101%	08:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	08:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WEIST

First Name: JAMES

MI: R

Signature: _____

See J AT 352

Badge No.: 352

Date: 01/28/2022

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C	Serial No.: ARUM-0043
Location: CHERRY HILL TWSP. P.D.	
Calibration File No.: 02633	Calib. Date: 11/16/2021
Certification File No.: 02634	Calib. No.: 00042
Linearity File No.: 02635	Cert. Date: 11/16/2021
Solution File No.: 02654	Cert. No.: 00038
Sequential File No.: 02654	Lin. Date: 11/16/2021
	Lin. No.: 00039
	Soln. Date: 01/03/2022
	Soln. No.: 00363
	File Date: 01/03/2022
Calibrating Unit: WET	Model No.: CU-34
Control Solution %: 0.100%	Serial No.: DDWA S3-0008
Solution Control Lot: 21090	Expires: 03/17/2023
	Bottle No.: 0036

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:38S		
Control 1 EC	0.101%	09:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40S		
Control 2 EC	0.100%	09:40S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41S		
Control 3 EC	0.100%	09:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WEIST	First Name: JAMES	MI: R
Signature: <u>Sgt [Handwritten Signature] 352</u>	Badge No.: 352	Date: 01/03/2022