

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Cherry Hill Twp. Police Department 820 Mercer St Cherry Hill NJ 08002	2. Contact: PO James Watts	4. Date: 10/20/2016
	3. Phone Number: 856-488-7828	5. County: Camden
6. Alcotest Instrument Serial Number: ARUM-0043	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
The Alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:
See Draeger Return and Repair form.

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Matthew R. Watson #7078
Name & Badge Number (Print)

Tpr. II Matthew R. Watson #7078
Signature Date **10/20/16**

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Cherry Hill Twp. Police Department 820 Mercer St Cherry Hill NJ 08002	2. Contact: PO James Watts 3. Phone Number: 856-488-7828	4. Date: 09/01/2016 5. County: Camden
6. Alcotest Instrument Serial Number: ARUM-0043	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:
I responded to the Cherry Hill Police Department for a semi-annual calibration of their Alcotest. While on scene I received complaints stating that the Alcotest experienced Control Test Failures on three subsequent Breath Tests. (Sequential file numbers 1827, 1829 and 1833)


10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument via the data downloads and the aforementioned alcohol influence reports, Sequential file numbers 1827, 1829 and 1833. This revealed that the control test failures are due to results outside of tolerance. Instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 001870

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. II Matthew R. Watson #7078
Name & Badge Number (Print)


 Signature

 #7078
 9/1/16
 Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Cherry Hill Twp. P.D., Cherry Hill, NJ

Date Received: 09/06/2016

Date given to service: 09/06/2016

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 7110 8610 7510 DT5000 7410 Upper-half 9510 7410 Whole

Serial Number: AR UM-0043
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: December 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARJC-2228

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 10/07/2016

Service note

Customer no. 150054969 Number of report 306033800 Date of receiving 04/21/2021

Please reference on inquiries

Customer
 CHERRY HILL POLICE DEPARTMENT
 820 MERCER ST
 CHERRY HILL NJ 08002-2688

Consignee
 CHERRY HILL POLICE DEPARTMENT
 820 MERCER ST
 CHERRY HILL NJ 08002-2688

Your order

Date of order:
 04/21/2021
 Your reference:
 P1-08791

NL-Text

Maintenance device list

Your contact person

Service Support-IDT
 800-437-2437 OPT 5
 SS-IDT@draeger.com

Page 1

Item	Quantity	Part no. Description
		Service Order: 144715326 CALIBRATION/ DDWAS3-0007 Eq.No. : 1023340906 SN: DDWAS3-0007 Material: 4540083 Simulator,CU34 110V elec ctrl w/o tmeter Inv.No. : Location:
0.0	H	CALIBRATION/ DDWAS3-0007 R001 Repair time - normal hours
1	EA	4412034 Hose pump-sim w/sim-cuvette hose 2" KIT

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Cherry Hill Twsp. PD 820 Mercer Street Cherry Hill, NJ 08002	2. Contact: P/O J. Watts #351	4. Date: 1-13-2011
	3. Phone Number: 856-432-8859	5. County: Camden

6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARUM-0043</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>
--	--	--

9. Reason for Service:
The Alcotest 7110 came back from outside evaluation from Draeger and is returned to service.

10. Comments:
See attached packing slip for details of Draeger's work conducted on the instrument.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I F. Sullivan #5103	<i>Tpr. I F. Sullivan #5103</i>	1-13-2011
Name & Badge Number (Print)	Signature	Date

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Cherry Hill Twsp. Police Department 820 Mercer Street Cherry Hill, NJ 08002	2. Contact: P/O J. Watts #351 3. Phone Number: 856-432-8859	4. Date: 12-22-2010 5. County: Camden
6. Alcotest Instrument Serial Number: ARUM-0043	7. Simulator Component Serial Number: n/a	8. Temperature Probe Component Serial Number: n/a

9. Reason for Service:
 Fuel cell would not recovery after control tests.

10. Comments:
 Instrument was sent back to Draeger for further evaluation.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I F. Sullivan #5103
 Name & Badge Number (Print)

Tpr. I F. Sullivan #5103 12-22-10
 Signature Date